AMENDMENT NO. 1 SEPTEMBER 1980

TO

IS : 7115-1973 SPECIFICATION FOR APPARATUS, TUBAL PATENCY

Alteration

(*Page* 2, *Clause* **5.1**) — Substitute the following for the existing clause:

" **5.1** *Corrosion Resistance Test* (*Applicable Only to Stainless Steel Cannula*) — Test the cannula in accordance with IS : 7531-1975 ' Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments '. The cannula shall show no sign of corrosion after the test. "

(CPDC 25)

Printed at Arcee Press, New Delhi, India